

Gausul Azam BNSB Eye Hospital, Newtown, Dinajpur
Phone- 0531-64033

Subject: Invitation of quotation of Phaco Emulsification Machine.

This is to inform to all the concern of Ophthalmic Equipment & Instrument Importers/ Suppliers/ Vendors that Gausul Azam BNSB Eye Hospital has decided to purchase a latest model brand new Phaco Emulsification Machine for the Hospital. The Importers/ Suppliers/ Vendors who are covering the following specification and Accessories of your machine are requested to submit Quotation on or before 24 March, 2016. If any Vat & Tax is applicable to pay for this purpose the supplier/ vendor is responsible to pay this and after payment of Vat & Tax a Chalan copy will be submitted to the Hospital Authority. The Hospital authority keeps & reserves all rights to accept or reject any quotation and the whole process also at any time without assigning any reason. The Suppliers/ Vendors are requested to submit Quotations hard copy along with catalog and all details information.

SPECIFICATION:


- Console 8065750833 1EA 0
- Foot Switch 8065750403 1EA 0
- Remote Control 8065750468 1EA 0
- Ozil Phaco Handpiece 8065750469 1EA 0
- I/A HP Threaded STTL Ultraflow SP 806581 7801 1EA 0
- Intreoid I/A Tip 0.3 mm 8065751 01 3 1EA 0
- Infiniti Anterior VitrectomY Pak 80657501 57 1 Box 0
- Bipolar Cable 12 ft.Silicone 8065128402 1EA 0
- Biolar Forceps 80651 29301 1EA 0
- Memory Card 8065750254 1EA 0
- injector (Monarch III) 8065977773
- Fusion pump consisting peristaltic and Venturi Pump (Dual Pump)
- Transversal Ultrasound Emulsification Technology.
- Customized settings option by surgeons.
- Wireless remote Control
- Controlled via the front panel. footswitch & remote control
- Several modalities of ultrasonic power control including continuous hyper pulsed and micro “burst” application of ultrasonic power
- Voice confirmation during mode changes.
- Flat screen, active matrix- color LCD display with touch screen.
- Disposable Cassette.
- Sub Modes: Pulse, continuous Burst
- Tip Stock @ 100%:88.9
- Anterior Vitrectomy:
- Six-crystal handpiece with internal infusion line.

- Ultrasound Time: System records and displays ultrasound time in 0.01 second increments
- Programming Levels 1,2,3:Continuous ultrasound, Pulsed ultrasound, Fixed pulse ultrasound, Single, burst ultrasound, Multiple burst ultrasound (Range: 2 to 600 msec).
- Provides Cassette Full, Near-Full and Continuous Fluid Level Sensing.
- Linear control of vacuum, Fixed, On / Off control of vacuum, Dual Linear Modes: Pitch or Yaw.
- Vitrectomy - Linear Cut Rate Mode Operating Parameters Range: 0 to 800 cuts per minute in 1 cut per minute increments Control: Linear control of cut rate via the foot pedal.
- Foot pedal
- Bipolar (Coagulation)
- IV Pole
- Computer unit
- Incision size

ACCESSORIES:

- Dual line up Foot paddle
- Operator manual
- Remote control
- Dust cover
- H/P sterilization tray
- Phaco hand piece
- 1/A handle, Titanium
- 1/A Tip
- laminar flow phaco tip straight/ curved
- next generation phaco pack
- bipolar cord
- bipolar forceps
- laminar flow metal tip wrench

Thanking all



10-03-2016

Dr. Chow. Mosaddequl Isdani
 General Secretary
 Gausul Azam BNSB Eye Hospital, Dinajpur.